

Catell rekommenderar att produkten provas ut och anpassas av legitimerad eller övrigt utbildad sjukvårdskunnig personal. Brukare med diabetes, känslig hud, kända kontaktallergier, nedsatt känsel eller dålig cirkulation ska vara extra observant för tryck, skav eller ödem under användning. Kontrollera alltid materialinnehåll nedan. Alla allvariga incidenter i samband med användning av denna produkt ska anmälas till tillverkaren och till behörig myndighet i den medlemsstat där användaren och/eller patienten är etablerad. Om frågor kring produkten uppstår kontakta din ordinatör, utprovare eller Catell.

Användningsområde

Ortosen stödjer och avlastar berört finger efter skada, operation, inflammatorisk tillstånd, kronisk sjukdom eller överbelastningsbesvär.

Applicerings

1. Aluminiumramen i skenan anpassas efter brukarens finger (basfalang) samt önskad position. Bild 1
2. Placera ortosen på fingret. Bild 2
3. Stäng kardborrbandet så att ortosen sitter behagligt. Tänk på att inte dra åt för hårt.



Justera kardborrbanden vid behov. Det är viktigt att ortosen sitter bekvämt, utan att klämma, skava eller orsaka stas.

Material

Nylon, polyester, aluminiumram. Produkten är latexfri.

Tvättråd

Slut kardborrbanden. Följ tvättråden nedan. Separat tvätt rekommenderas. Använd gärna tvättpåse. Låt ortosen dropptorka.



Måttabell

Strl	*Mått
S	5 - 6,5 cm
M	6,5 - 8 cm
L	8 - 9 cm

*Mått tas cirkulärt runt proximala falangen och ska ses som en vägledning.

Produkten är bilateral.

Denna produkt är endast avsedd att användas på en patient.

Återvinning

Sortera enligt lokala föreskrifter.

Denna produkt överensstämmer med de krav som ställs i den Medicintekniska förordningen (EU) 2017/745, klass I produkter.



Manufacturer Catell AB

Västberga Allé 26, SE-126 30 Hägersten, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

Instruction for use – Triggerfix



Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals. Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the brace. Always check material content below. Any serious incidents occurring relates to the device should be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is resident. Contact your prescriber or Catell if questions about the product occur.

Area of use

The brace supports the users finger after injury, surgery, inflammatory conditions, chronic disease or overuse problems.

Application

1. Adjust the aluminiumbar in the brace to the user's finger (proximal phalanx) and desired position.
Picture 1
2. Place the brace on the finger. Picture 2
3. Close hook and loop strap so that the brace fits comfortably. Remember not to tighten too much.



Adjust the hook and loop straps when necessary. It is important that the brace fits comfortably without pinching, chafing or affect blood circulation.

Material

Nylon, polyester, aluminiumbar. Free from latex.

Washing instructions

Close the hook and loop straps. Follow the washing instructions below. Separate wash is recommended. Use a washing pouch. Let the brace drip dry.



Size

Size	*Dimensions
S	5 - 6,5 cm
M	6,5 - 8 cm
L	8 - 9 cm

*Measure proximal phalanx circumference. Measurements should be seen as guidelines.

Bilateral.

Single patient use.

Recycling

Sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation (EU) 2017/745 product class I.



Manufacturer Catell AB
Västberga Allé 26, SE-126 30 Hägersten, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

220701:2